

Amendments to the Specification

(N4) Please replace the table that follows the title "Table 1 – Comparison of Expansile Forces of a Z-Stent, a WALLSTENT and a Nitinol Woven Stent" on page 30 with the following table:

Δ (mm)	Z Center (g)	Z Between (g)	Z Side by Side (g)	W Center (g)	W Overlap (g)	W Side by Side (g)	Woven Stent (g)
2	16	13	19	15	35	18	44
4	36	28	31	25	59	22	91
6	51	44	42	42	80	35	126
8	63	61	56	50	108	42	158
10	81	79	62	60	126	48	167
12	100	98	76	74	149	54	175
14	115	119	90	84	170	63	184
16	127	133	101	100	197	73	202
18	146	192	122	111	220	84	
20	165	unmeasur.	142	129	248	96	